

Florin Analytical Services (FAS), 2016 Society of Mineral Analyst (SMA) Round Robin Results

Sample		1A	1B	2A	2B	3A	3B
FAS Results	Mar-16	0.0230	0.0228	0.0092	0.0091	81.1762	80.7997
32 Participating Laboratory	Mean	0.0244	0.0244	0.0093	0.0093	73.8832	73.8832
32 Participating Laboratory	95% Confidence	0.0070	0.0070	0.0006	0.0006	3.2691	3.2691
FAS Results	May-16	0.0373	0.0373	0.3148	0.3143	4.5541	4.5134
30 Participating Laboratory	Mean	0.0365	0.0365	0.3007	0.3007	4.3268	4.3268
30 Participating Laboratory	95% Confidence	0.0007	0.0007	0.0106	0.0106	0.0994	0.0994
FAS Results	Jul-16	0.0936	0.0937	0.0444	0.0442	116.1330	115.6120
29 Participating Laboratory	Mean	0.0925	0.0925	0.045	0.045	114.4447	114.4447
29 Participating Laboratory	95% Confidence	0.0014	0.0014	0.0015	0.0015	2.3562	2.3562
FAS Results	Sep-16	0.0279	0.0277	0.0053	0.0055	88.3243	88.6979
29 Participating Laboratory	Mean	0.0325	0.0325	0.012	0.012	84.393	84.393
29 Participating Laboratory	95% Confidence	0.0019	0.0019	0.0027	0.0027	0.9335	0.9335
FAS Results	Nov-16	0.0328	0.0327	0.2031	0.2004	8.5940	8.4910
26 Participating Laboratory	Mean	0.0335	0.0335	0.2055	0.2055	8.4049	8.4049
26 Participating Laboratory	95% Confidence	0.0006	0.0006	0.0041	0.0041	0.0738	0.0738

